



Calibration Certificate

Product Code:
DMM-8062A

Description:
Personnel Tester

Serial Number: _____

- Threshold settings prior to calibration:

This is a new calibration there are no previous calibration values.

- Calibration of this instrument is hereby certified to be within the published specification as shown below:

DCV	5.5V	5.476~5.524V
ACV	5.5V/60Hz	5.454~5.546V
Ω	550 Ω	544.6~555.4 Ω
DCA	20A	19.77~20.23A
ACA	20A/60Hz	19.64~20.36A
CAP	55nF	53.64~56.36nF
Hz	10Hz/200mV rms	9.96~10.04Hz
$^{\circ}\text{C}$	1000 $^{\circ}\text{C}$	990~1010 $^{\circ}\text{C}$

- The instrument is calibrated against standards traceable to **CE** standards.
- Details of reference equipment used:
Calibration FLUKE 5522A Serial Number: 3917903
- Certificate of reference equipment:
Issue Date: Certificate Number: 202401000183

Equipment Control Number:

Date:

Signed:

CTS/T/5522A/002

Name:

Title: Quality Supervisor